



MANUFACTURER'S DECLARATION: MEDICAL DEVICES DIRECTIVE 93/42/EEC

As this component is not a complete "Medical Device" under the definition provided in the Directive but a part with no function on its own, a Declaration of Conformity or CE mark is not appropriate until the component is incorporated. The following Declaration is issued in order to aid the installation of the component below into a finished appliance, for use by, or sale to, an end-user.

The following declares the status of these products with regard to Union Directive 93/42/EEC:

Product Name	Product Number(s)
VSi	D50242, D50464, D50514, D50929, D50963, D50979, D51049, D51112, D51250, D51409, D50172, D50301, D50416, D51161, D51162, D51542

The object of the declaration described above is in conformity with the relevant Union harmonisation legislation to the extent possible as a component to be integrated into a complete product.

HARMONISED STANDARDS

The following harmonised standards (and those called by them) were used in order to attain a presumption of conformity with the essential requirements of the above directive as far as the component allows:

- EN 12184:2009 Electrically powered wheelchairs, scooters and their chargers - Requirements and test methods
- EN 14179:2009 Medical devices — Application of risk management to medical devices
- EN 62304:2006 + A1:2015 Medical device software — Software life-cycle processes
- EN 12182:2012 Assistive products for persons with disability. General requirements and test methods

(Detail of Essential Requirement Coverage overleaf)

Nigel Sharp, Senior Consultant Engineer

12 March 2020

Signed at, for and on behalf of:

Penny & Giles Controls Ltd.,
15 Enterprise Way,
Aviation Park West,
Christchurch,
Dorset, UK
BH23 6HH